



REPORT
Intertek Testing Services, NA Inc.
1717 Arlingate Lane COLUMBUS, OHIO 43228

PROJECT NO.: G100385167

DATE: April 29, 2011

REPORT NO. 100385167COL-001

RENDERED TO:
Blue Wave International Corp
1867 Ridgecrest Rd
Prescott, AZ 86305

STANDARD REFERENCED AND TEST METHOD:

ITS Non-Standardized Test: Microbial Reduction Rate Test

AUTHORIZATION:

The test was authorized by Rolf Engelhard; A representative from Blue Wave International Corp.

GENERAL DESCRIPTION: The test performed was the Microbial Reduction Rate Test conducted at the Intertek microbiology lab in Columbus, Ohio. The models Air Purifier #1 and Air Purifier #2 were tested for their ability to reduce the number of microorganisms aspirated into a 411.4 cu ft test room. The sample was received on April 14, 2011 and they are currently prototype models. The test chamber was contaminated with *Penicillium citrinum* (ATCC#36382) and *Staphylococcus epidermis* (ATCC# 12228).

TEST DESCRIPTION

Nutrient agar was prepared for the bacteria cultures and potato dextrose agar was prepared for the mold cultures.

All agars were sterilized using an autoclave to a temperature of 121°C.

The bacterial and mold cultures were prepared using pre-grown cultures acquired from ATCC (American Type Culture Collection with respective numbers describing each microorganism).

Using an inoculating loop, the cultures were transferred daily in nutrient/yeast broth for not more than two weeks. Nutrient broth was used for the growth of bacteria. Yeast broth was used for the growth of mold. At the conclusion of two weeks, a fresh transplant from stock culture was made. Bacterial cultures were incubated at $37 \pm 2^\circ\text{C}$ for 24 hours. Mold cultures were grown at 30-32°C and 85 % relative humidity for 28 days.

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.

REPORT NO.:100385167COL-001

DATE: April 29, 2011

The stock cultures were maintained on nutrient agar. Cultures were stored at $5 \pm 1^{\circ}\text{C}$ and transfer once a month.

This operation was completed for each microorganism.

Samples were set in the center of the testing room. Activation was performed by manipulation of the power source from outside the chamber.

The collision nebulizer was then put into the test chamber where it was attached to an Erlenmeyer vacuum flask and a nitrogen tank. The nozzle of the flask pointed outward toward the room.

The room (411.4 cubic feet) was now sealed and a negative control was taken. This ensures that there were no other microorganisms in the test chamber prior to testing.

The nitrogen for the aspirator was set and started for aspiration into the test chamber.

A positive control sample of the air was now to be taken. This provided reaffirming data that the correct amount of the microorganism was put into the test chamber.

Samples were taken every 30 minutes from the air sampler that was attached to the chamber wall. The agar plates were put into the air sampler and the microorganism was vacuumed onto the plate.

The bacteria samples were then put into the incubator at their appropriate temperatures and allowed to grow for 48 hours. Mold cultures were grown in the environmental chamber at $30\text{-}32^{\circ}\text{C}$ and 85 % relative humidity for 2-3 days.

This process was repeated as above, this time each air cleaner was turned on at time zero or when you take the first sample. These results were then compared to the natural decay of the microorganism to arrive at percent reduction. Each air purifier was allowed 10 minutes to warm up per the manufacturer's instructions.

This process was performed for each microorganism.

CALIBRATED EQUIPMENT:

Calibrated Equipment	Manufacturer	Identification No.	Calibration Date	Calibration Due
Micropipette	Fisher Scientific	CE 1141	03/12/11	03/12/12
Incubator	Precision	CE 1155	03/12/11	03/12/12
Thermometer	Precision	CE 1135	03/12/11	03/12/12
Environmental Chamber	Thermotron	CE 1142	For Reference only	See CE 1151
Digital Hygrometer	Fisher Scientific	CE 1151	01/12/11	01/12/12

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.

REPORT NO.:100385167COL-001

DATE: April 29, 2011

RESULTS:

The negative controls showed no signs of growth.

The positive controls showed complete growth over the agars surface. The original number of each microorganism aspirated into the chamber was 1×10^8 cfu/ml.

Air Purifier #1-

S. epidermis has shown;

- A 50.6% reduction from the natural decay after 1 hour,
- A 59.4% reduction from the natural decay after 2 hours,
- A 93.4% reduction from the natural decay after 4 hours

P. citrinum has shown;

- A 47.4% reduction from the natural decay after 1 hour,
- A 66.3% reduction from the natural decay after 2 hours,
- A 92.7% reduction from the natural decay after 4 hours

Air Purifier #2-

P. citrinum has shown;

- A 50.0% reduction from the natural decay after 1 hour,
- A 64.0% reduction from the natural decay after 2 hours,
- A 90.2% reduction from the natural decay after 4 hours

Penicillium citrinum

Time (minutes)	Natural Decay (colonies) CFU	Air Purifier #1 (colonies) CFU	Air Purifier #2 (colonies) CFU
0	138	139	136
30	129	101	108
60	116	61	58
90	98	38	42
120	86	29	31
150	69	23	29
180	54	17	22
210	47	11	13
240	41	3	4
Total End % Reduction	N/A	92.7%	90.2%

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.

REPORT NO.:100385167COL-001

DATE: April 29, 2011

Staphylococcus epidermis

Time (minutes)	Natural Decay (colonies) CFU	Air Purifier #1 (colonies) CFU
0	287	285
30	252	174
60	241	119
90	229	91
120	212	86
150	189	63
180	161	33
210	126	17
240	91	6
Total End % Reduction	N/A	93.4%

CONCLUSION: This report documents the performance of the air purifier #1 and air purifier #2. The microbiological test sample evaluations were conducted at the Intertek laboratory located in Columbus, OH between April 18, 2011 and April 29, 2011. The air purifier #1 does successfully reduce *S. epidermis* 93.4% and *P. citrinum* 92.7% in 4 hours. The air purifier #2 does successfully reduce *P. citrinum* 90.2% in 4 hours.

Test Performed by:



Shannon Meier
Microbiologist
Columbus Office

Report Approved by:



Ramzi Amawi
Operations Manager
Columbus Office

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.